
Midmark M11 - v1.0.0
Total Cycles: 2835

Name: RITTER 1
Sterilizer ID: V2178198
01/31/2023 02:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 210.0 F 0.0
2:00 203.4 F 4.7
4:00 211.7 F 9.9
6:00 224.6 F 6.3
8:00 234.3 F 10.9
10:00 235.8 F 8.1
12:00 241.0 F 10.7
14:00 247.1 F 13.7
16:00 259.7 F 20.6
18:00 271.1 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.2
1:00 272.4 F 28.9
2:00 271.6 F 28.3
3:00 272.0 F 28.9
4:00 272.2 F 28.8

Min 271.1 F 28.1
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.7 F 0.0
5:00 223.2 F 0.0
10:00 230.7 F 0.0
15:00 231.4 F 0.0
20:00 225.6 F 0.0
25:00 225.2 F 0.0
30:00 224.8 F 0.0
Drying Complete

FILLING: 2:09
HEATING: 18:44
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:58:12
CYCLE COMPLETE

01/31/2023 03:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____