
Midmark M11 - v1.0.5
Total Cycles: 2599

Name: RITTER2
Sterilizer ID: V2313228
07/04/2023 08:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.3 F 0.0
2:00 173.4 F 0.5
4:00 170.9 F 3.9
6:00 194.9 F 8.1
8:00 214.2 F 13.1
10:00 225.9 F 8.3
12:00 237.3 F 13.3
14:00 237.6 F 10.5
16:00 236.9 F 8.5
18:00 248.7 F 15.1
20:00 260.1 F 21.5
22:00 270.4 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.1
1:00 272.2 F 29.9
2:00 272.6 F 29.8
3:00 272.0 F 30.0
4:00 272.7 F 29.9

Min 271.0 F 28.9
Max 273.1 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.2 F 0.0
5:00 218.5 F 0.0
10:00 221.9 F 0.0
15:00 226.1 F 0.0
20:00 227.2 F 0.0
25:00 227.8 F 0.0
30:00 223.3 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 22:55
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:01:23
CYCLE COMPLETE

07/04/2023 10:00 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____