
Midmark M11 - v1.0.5
Total Cycles: 1947

Name: RITTER2
Sterilizer ID: V2313228
01/19/2023 08:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.1 F 0.0
2:00 78.9 F 0.0
4:00 137.1 F 0.0
6:00 170.9 F 0.0
8:00 197.7 F 0.1
10:00 191.7 F 0.3
12:00 206.2 F 3.1
14:00 217.1 F 6.3
16:00 228.1 F 9.9
18:00 234.0 F 10.7
20:00 238.5 F 10.6
22:00 235.5 F 8.4
24:00 242.0 F 11.0
26:00 247.1 F 14.0
28:00 259.3 F 20.7
30:00 270.3 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.4 F 29.1
2:00 271.7 F 28.9
3:00 272.3 F 29.0
4:00 271.7 F 28.9

Min 271.5 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.4 F 0.0
5:00 218.6 F 0.0
10:00 195.0 F 0.0
15:00 196.7 F 0.0
20:00 214.1 F 0.0
25:00 228.7 F 0.0
30:00 217.1 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 30:50
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:08:46
CYCLE COMPLETE

01/19/2023 09:16 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____