
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000296
10:36 05/08/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:06
STERILIZING 4:02
136.0 C 220kPa 4:02
136.4 C 223kPa 4:32
136.0 C 222kPa 5:02
136.2 C 224kPa 5:32
136.2 C 218kPa 6:02
136.2 C 223kPa 6:32
136.4 C 223kPa 7:02
136.4 C 224kPa 7:32
136.4 C 224kPa 7:32
VENTING 7:32
AIR DRYING 8:22
CYCLE COMPLETE 8:29

Digital Signature #
8AB5C4379A4F8353
