
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000052
15:02 21/09/2020
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:58
PRESSURIZING 8:12
STERILIZING 10:00
136.0 C 223kPa 10:00
136.0 C 225kPa 10:30
136.4 C 227kPa 11:00
136.3 C 224kPa 11:31
136.4 C 226kPa 12:01
136.4 C 227kPa 12:31
136.4 C 226kPa 13:01
136.5 C 226kPa 13:30
VENTING 13:30
AIR DRYING 14:53
CYCLE COMPLETE 44:53

Digital Signature #
B13141D35AAC349A
