
Midmark M11 - v1.0.5
Total Cycles: 3894

Name: RITTER2
Sterilizer ID: V2313228
08/27/2024 11:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.2 F 0.0
2:00 70.9 F 0.0
4:00 104.2 F 0.0
6:00 111.6 F 0.1
8:00 142.7 F 0.6
10:00 171.3 F 3.6
12:00 187.7 F 7.3
14:00 204.2 F 11.4
16:00 214.6 F 4.6
18:00 226.8 F 8.7
20:00 237.6 F 13.0
22:00 237.7 F 10.2
24:00 236.4 F 8.7
26:00 248.8 F 14.9
28:00 260.3 F 21.3
30:00 270.8 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.3 F 28.9
2:00 272.7 F 29.4
3:00 271.8 F 28.9
4:00 272.4 F 28.9

Min 271.5 F 28.5
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.7 F 0.0
5:00 199.5 F 0.0
10:00 195.3 F 0.0
15:00 193.6 F 0.0
20:00 207.3 F 0.0
25:00 216.5 F 0.0
30:00 233.2 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 30:42
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:08:43
CYCLE COMPLETE

08/27/2024 12:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____