
Midmark M11 - v1.0.5
Total Cycles: 2658

Name: RITTER2
Sterilizer ID: V2313228
07/17/2023 05:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.6 F 0.0
2:00 159.0 F 2.1
4:00 183.5 F 6.5
6:00 209.9 F 12.1
8:00 223.0 F 8.1
10:00 233.8 F 10.6
12:00 240.1 F 12.2
14:00 240.0 F 11.1
16:00 253.8 F 18.1
18:00 265.8 F 25.3
20:00 272.8 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 29.7
1:00 271.9 F 29.9
2:00 272.4 F 30.1
3:00 272.1 F 29.9
4:00 271.7 F 29.8

Min 271.6 F 29.6
Max 272.7 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.4 F 0.0
5:00 220.9 F 0.0
10:00 217.1 F 0.0
15:00 219.9 F 0.0
20:00 216.2 F 0.0
25:00 216.1 F 0.0
30:00 211.2 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 20:02
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:58:48
CYCLE COMPLETE

07/17/2023 06:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____