

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3983

Name: RITTER 1  
Sterilizer ID: V2178198  
11/06/2023 11:31 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 193.7 F 0.0  
2:00 178.3 F 2.7  
4:00 186.9 F 6.3  
6:00 222.6 F 10.5  
8:00 223.7 F 6.3  
10:00 226.1 F 8.6  
12:00 240.3 F 13.1  
14:00 237.2 F 9.8  
16:00 238.6 F 9.5  
18:00 247.0 F 13.6  
20:00 258.0 F 19.5  
22:00 267.9 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.1  
1:00 272.5 F 28.9  
2:00 272.1 F 28.4  
3:00 271.2 F 28.2  
4:00 272.2 F 28.9

Min 271.2 F 28.0  
Max 272.6 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.3 F 0.0  
5:00 227.5 F 0.0  
10:00 222.7 F 0.0  
15:00 228.0 F 0.0  
20:00 228.3 F 0.0  
25:00 222.5 F 0.0  
30:00 226.6 F 0.0  
Drying Complete

FILLING: 1:38  
HEATING: 23:26  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:02:47  
CYCLE COMPLETE

11/06/2023 12:33 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_