
Midmark M11 - v1.0.5
Total Cycles: 4511

Name: RITTER2
Sterilizer ID: V2313228
04/14/2025 04:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 90.3 F 0.0
2:00 89.5 F 0.3
4:00 100.0 F 1.8
6:00 139.1 F 4.3
8:00 152.6 F 7.4
10:00 177.6 F 10.8
12:00 217.1 F 7.7
14:00 215.7 F 6.2
16:00 225.3 F 9.9
18:00 231.9 F 10.4
20:00 236.3 F 10.3
22:00 236.8 F 9.3
24:00 246.3 F 14.0
26:00 256.8 F 19.7
28:00 266.8 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.3
1:00 272.7 F 30.0
2:00 272.5 F 30.0
3:00 272.6 F 30.1
4:00 272.5 F 30.2

Min 271.0 F 29.3
Max 273.3 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.2 F 0.0
5:00 235.7 F 0.0
10:00 216.2 F 0.0
15:00 218.3 F 0.0
20:00 219.0 F 0.0
25:00 221.8 F 0.0
30:00 220.1 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 29:37
STERILIZING: 4:00
VENTING: 3:54
DRYING: 30:00
TOTAL CYCLE: 01:08:24
CYCLE COMPLETE

04/15/2025 12:44 PM
ACCEPTED: X
REJECTED:
Operator: U01