

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2721

Name: RITTER 1  
Sterilizer ID: V2178198  
12/22/2022 03:07 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 197.6 F 0.0  
2:00 196.0 F 3.8  
4:00 209.1 F 8.9  
6:00 224.9 F 7.3  
8:00 231.9 F 10.0  
10:00 239.7 F 8.2  
12:00 244.8 F 12.7  
14:00 244.1 F 12.6  
16:00 257.3 F 19.6  
18:00 269.0 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.6  
1:00 272.6 F 29.0  
2:00 272.5 F 29.2  
3:00 272.4 F 28.8  
4:00 272.6 F 29.4  
  
Min 271.8 F 28.5  
Max 272.9 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.9 F 0.0  
5:00 231.9 F 0.0  
10:00 201.6 F 0.0  
15:00 204.8 F 0.0  
20:00 241.3 F 0.0  
25:00 227.5 F 0.0  
30:00 227.7 F 0.0  
Drying Complete

FILLING: 2:30  
HEATING: 19:03  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:58:46  
CYCLE COMPLETE

12/22/2022 04:05 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_