

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 008370  
14:43 27/02/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:38  
PRESSURIZING 2:43  
STERILIZING 3:51  
136.4 C 224kPa 3:51  
136.3 C 224kPa 4:21  
136.4 C 224kPa 4:51  
136.3 C 224kPa 5:21  
136.2 C 221kPa 5:51  
136.4 C 224kPa 6:21  
136.3 C 224kPa 6:51  
136.1 C 219kPa 7:21  
136.1 C 219kPa 7:21  
VENTING 7:21  
AIR DRYING 8:10  
CYCLE COMPLETE 38:10

-----  
Digital Signature #  
5BAA5C619F380C5C  
-----