
Midmark M11 - v1.0.0
Total Cycles: 645

Name: MIDMARK
Sterilizer ID: V2178198
02/01/2021 11:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 137.4 F 0.0
2:00 136.8 F 0.8
4:00 170.1 F 3.6
6:00 180.6 F 7.9
8:00 206.2 F 13.3
10:00 216.8 F 7.4
12:00 231.3 F 12.9
14:00 238.2 F 10.0
16:00 235.0 F 7.7
18:00 250.2 F 15.1
20:00 262.7 F 22.2
22:00 271.6 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.1
1:00 271.8 F 28.5
2:00 272.0 F 29.0
3:00 272.7 F 29.0
4:00 272.5 F 28.9

Min 271.0 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.3 F 0.0
5:00 235.6 F 0.0
10:00 203.1 F 0.0
15:00 201.0 F 0.0
20:00 200.0 F 0.0
25:00 198.9 F 0.0
30:00 198.2 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 22:05
STERILIZING: 4:00
VENTING: 2:56
DRYING: 30:00
TOTAL CYCLE: 00:59:59
CYCLE COMPLETE

02/01/2021 12:28 PM
ACCEPTED: X
REJECTED:
Operator: U01