
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000419
9:35 27/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:11
PRESSURIZING 2:18
STERILIZING 3:23
136.0 C 220kPa 3:23
136.4 C 228kPa 3:53
136.3 C 228kPa 4:23
136.3 C 231kPa 4:53
136.0 C 229kPa 5:23
136.4 C 232kPa 5:53
136.3 C 230kPa 6:23
136.3 C 233kPa 6:53
136.3 C 233kPa 6:53
VENTING 6:53
AIR DRYING 7:45
CYCLE COMPLETE 7:53

Digital Signature #
0E4A0923159F2ED2
