
Midmark M11 - v1.0.0
Total Cycles: 1515

Name: RITTER 1
Sterilizer ID: V2178198
01/12/2022 12:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.3 F 0.0
2:00 197.6 F 3.8
4:00 215.1 F 8.9
6:00 221.7 F 7.5
8:00 232.4 F 10.0
10:00 236.6 F 8.2
12:00 244.5 F 12.7
14:00 244.6 F 12.6
16:00 257.6 F 19.5
18:00 269.2 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.3 F 28.6
2:00 272.5 F 29.1
3:00 271.6 F 28.3
4:00 272.8 F 29.1

Min 271.5 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.4 F 0.0
5:00 212.8 F 0.0
10:00 224.2 F 0.0
15:00 227.7 F 0.0
20:00 227.4 F 0.0
25:00 226.5 F 0.0
30:00 224.8 F 0.0
Drying Complete

FILLING: 2:18
HEATING: 18:59
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:29
CYCLE COMPLETE

01/12/2022 01:23 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____