

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4508

Name: RITTER 1  
Sterilizer ID: V2178198  
05/01/2024 09:23 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.9 F 0.0  
2:00 175.8 F 2.5  
4:00 178.0 F 6.7  
6:00 194.2 F 11.2  
8:00 214.6 F 4.9  
10:00 226.0 F 10.0  
12:00 232.0 F 8.3  
14:00 241.3 F 12.2  
16:00 240.0 F 10.7  
18:00 253.5 F 17.2  
20:00 264.9 F 24.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.5  
1:00 272.4 F 29.1  
2:00 271.4 F 28.6  
3:00 272.5 F 29.2  
4:00 271.6 F 28.5  
  
Min 271.2 F 28.4  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.2 F 0.0  
5:00 213.8 F 0.0  
10:00 223.5 F 0.0  
15:00 228.6 F 0.0  
20:00 230.8 F 0.0  
25:00 230.6 F 0.0  
30:00 229.5 F 0.0  
Drying Complete

FILLING: 1:14  
HEATING: 21:52  
STERILIZING: 4:00  
VENTING: 3:32  
DRYING: 30:00  
TOTAL CYCLE: 01:00:37  
CYCLE COMPLETE

05/01/2024 10:23 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_