
Midmark M11 - v1.0.5
Total Cycles: 3077

Name: RITTER2
Sterilizer ID: V2313228
11/02/2023 11:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.4 F 0.0
2:00 143.0 F 1.4
4:00 159.5 F 4.4
6:00 185.8 F 8.5
8:00 206.0 F 13.2
10:00 226.0 F 7.5
12:00 232.3 F 12.4
14:00 234.9 F 9.1
16:00 238.6 F 10.0
18:00 247.1 F 14.1
20:00 259.5 F 20.9
22:00 270.3 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 271.9 F 29.2
2:00 272.3 F 29.4
3:00 272.8 F 29.8
4:00 271.8 F 29.7

Min 271.5 F 29.2
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.7 F 0.0
5:00 194.9 F 0.0
10:00 194.9 F 0.0
15:00 195.8 F 0.0
20:00 194.6 F 0.0
25:00 194.2 F 0.0
30:00 192.0 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 23:25
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:01:52
CYCLE COMPLETE

11/02/2023 12:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____