

-----  
Midmark M11 - v1.0.5  
Total Cycles: 118

Name: RITTER2  
Sterilizer ID: V2313228  
06/18/2021 12:10 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 133.0 F 0.0  
2:00 134.9 F 1.8  
4:00 146.2 F 4.7  
6:00 171.3 F 8.7  
8:00 196.8 F 11.1  
10:00 212.9 F 6.9  
12:00 226.9 F 11.7  
14:00 231.1 F 8.9  
16:00 238.7 F 10.7  
18:00 245.0 F 13.9  
20:00 257.5 F 20.4  
22:00 268.5 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.5  
1:00 272.4 F 30.0  
2:00 271.9 F 30.0  
3:00 272.0 F 29.7  
4:00 272.1 F 30.2

Min 271.5 F 29.4  
Max 272.6 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.7 F 0.0  
5:00 211.9 F 0.0  
10:00 216.2 F 0.0  
15:00 217.4 F 0.0  
20:00 214.0 F 0.0  
25:00 212.2 F 0.0  
30:00 210.9 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 23:11  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:01:28  
CYCLE COMPLETE

06/18/2021 01:11 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_