
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001203
12:16 03/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:43
PRESSURIZING 3:44
STERILIZING 4:58
136.0 C 221kPa 4:58
136.3 C 225kPa 5:28
136.4 C 225kPa 5:58
136.3 C 224kPa 6:28
136.4 C 226kPa 6:58
136.3 C 224kPa 7:28
136.2 C 223kPa 7:58
136.4 C 225kPa 8:28
136.4 C 225kPa 8:28
VENTING 8:28
AIR DRYING 9:26
CYCLE COMPLETE 24:26

Digital Signature #
812E32AC14BD409B
