

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3823

Name: RITTER 1  
Sterilizer ID: V2178198  
09/26/2023 03:22 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.6 F 0.0  
2:00 178.2 F 2.5  
4:00 193.8 F 6.6  
6:00 211.7 F 11.5  
8:00 222.0 F 6.1  
10:00 230.4 F 11.0  
12:00 232.1 F 7.6  
14:00 244.7 F 13.0  
16:00 244.4 F 12.4  
18:00 256.4 F 18.6  
20:00 267.7 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.1 F 28.5  
2:00 272.2 F 28.9  
3:00 271.4 F 28.1  
4:00 272.2 F 28.8

Min 271.2 F 28.1  
Max 272.5 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.0 F 0.0  
5:00 223.0 F 0.0  
10:00 234.3 F 0.0  
15:00 230.6 F 0.0  
20:00 228.4 F 0.0  
25:00 226.0 F 0.0  
30:00 224.8 F 0.0  
Drying Complete

FILLING: 1:39  
HEATING: 21:18  
STERILIZING: 4:00  
VENTING: 3:44  
DRYING: 30:00  
TOTAL CYCLE: 01:00:41  
CYCLE COMPLETE

09/26/2023 04:22 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_