
Midmark M11 - v1.0.0
Total Cycles: 4063

Name: RITTER 1
Sterilizer ID: V2178198
11/29/2023 09:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 124.0 F 0.0
2:00 121.5 F 1.0
4:00 129.8 F 3.2
6:00 150.8 F 6.2
8:00 174.6 F 9.9
10:00 209.0 F 7.8
12:00 212.5 F 6.7
14:00 225.2 F 10.8
16:00 228.9 F 6.7
18:00 236.8 F 11.7
20:00 236.4 F 9.2
22:00 248.8 F 15.1
24:00 259.9 F 21.2
26:00 270.3 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.7
1:00 272.7 F 29.6
2:00 272.4 F 29.2
3:00 271.3 F 28.7
4:00 271.2 F 29.0

Min 271.0 F 28.6
Max 273.3 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.9 F 0.0
5:00 222.3 F 0.0
10:00 216.9 F 0.0
15:00 220.4 F 0.0
20:00 204.5 F 0.0
25:00 201.0 F 0.0
30:00 191.4 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 26:59
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:05:29
CYCLE COMPLETE

11/29/2023 10:59 AM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____