

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5269

Name: RITTER 1  
Sterilizer ID: V2178198  
01/23/2025 03:20 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 166.7 F 0.0  
2:00 170.9 F 3.7  
4:00 186.9 F 7.8  
6:00 203.5 F 12.2  
8:00 221.8 F 5.5  
10:00 230.0 F 10.1  
12:00 233.3 F 8.1  
14:00 242.2 F 11.8  
16:00 239.1 F 10.0  
18:00 252.1 F 16.3  
20:00 263.1 F 22.7  
22:00 272.8 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.3  
1:00 272.4 F 29.2  
2:00 272.6 F 29.0  
3:00 271.7 F 28.8  
4:00 272.7 F 29.4

Min 270.9 F 28.1  
Max 273.0 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.3 F 0.0  
5:00 204.3 F 0.0  
10:00 192.9 F 0.0  
15:00 193.5 F 0.0  
20:00 216.8 F 0.0  
25:00 238.2 F 0.0  
30:00 240.6 F 0.0  
Drying Complete

FILLING: 1:15  
HEATING: 22:19  
STERILIZING: 4:00  
VENTING: 3:48  
DRYING: 30:00  
TOTAL CYCLE: 01:01:22  
CYCLE COMPLETE

01/23/2025 04:21 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_