
Midmark M11 - v1.0.5
Total Cycles: 4488

Name: RITTER2
Sterilizer ID: V2313228
04/04/2025 12:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.9 F 0.0
2:00 149.8 F 2.2
4:00 158.9 F 5.2
6:00 184.0 F 9.2
8:00 204.7 F 10.8
10:00 220.2 F 7.3
12:00 228.1 F 12.2
14:00 232.2 F 9.1
16:00 238.5 F 10.2
18:00 245.5 F 13.6
20:00 257.4 F 20.0
22:00 268.1 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.5 F 29.7
2:00 271.1 F 29.2
3:00 272.4 F 29.8
4:00 271.3 F 29.1

Min 271.1 F 29.0
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.1 F 0.0
5:00 210.4 F 0.0
10:00 223.2 F 0.0
15:00 228.5 F 0.0
20:00 226.0 F 0.0
25:00 230.2 F 0.0
30:00 226.7 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 23:17
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:01:56
CYCLE COMPLETE

04/04/2025 01:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____