
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004287
13:41 02/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:52
PRESSURIZING 2:55
STERILIZING 3:56
136.4 C 221kPa 3:56
136.3 C 224kPa 4:27
135.9 C 220kPa 4:57
136.2 C 224kPa 5:27
136.3 C 223kPa 5:57
136.3 C 223kPa 6:27
136.2 C 220kPa 6:57
136.4 C 223kPa 7:27
136.4 C 223kPa 7:27
VENTING 7:27
AIR DRYING 8:55
CYCLE COMPLETE 10:30

Digital Signature #
996A9801E7BD367A
