

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1749

Name: RITTER 1  
Sterilizer ID: V2178198  
03/25/2022 02:03 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 182.6 F 0.0  
2:00 179.3 F 3.2  
4:00 196.7 F 7.1  
6:00 212.7 F 12.8  
8:00 223.9 F 6.9  
10:00 237.4 F 12.0  
12:00 234.8 F 9.1  
14:00 239.0 F 9.7  
16:00 247.1 F 14.0  
18:00 259.3 F 20.7  
20:00 270.4 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.6  
1:00 272.8 F 29.3  
2:00 271.7 F 28.5  
3:00 272.3 F 29.4  
4:00 272.5 F 29.0  
  
Min 271.3 F 28.4  
Max 272.8 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.8 F 0.0  
5:00 238.5 F 0.0  
10:00 228.8 F 0.0  
15:00 230.4 F 0.0  
20:00 224.3 F 0.0  
25:00 225.2 F 0.0  
30:00 219.6 F 0.0  
Drying Complete

FILLING: 1:39  
HEATING: 20:50  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:41  
CYCLE COMPLETE

03/25/2022 03:02 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_