
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000286
9:02 04/08/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:46
PRESSURIZING 3:47
STERILIZING 4:49
136.0 C 220kPa 4:49
136.3 C 219kPa 5:19
136.4 C 219kPa 5:49
136.3 C 222kPa 6:19
136.1 C 223kPa 6:49
136.3 C 223kPa 7:19
136.4 C 225kPa 7:49
136.1 C 221kPa 8:19
136.1 C 221kPa 8:19
VENTING 8:19
AIR DRYING 9:11
CYCLE COMPLETE 9:19

Digital Signature #
EBD77D885DAE9694
