
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 002671
10:08 11/11/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:30
PRESSURIZING 2:32
STERILIZING 3:32
136.4 C 223kPa 3:32
136.4 C 223kPa 4:02
136.2 C 223kPa 4:32
136.0 C 221kPa 5:02
136.4 C 222kPa 5:32
136.3 C 223kPa 6:02
136.3 C 223kPa 6:32
136.4 C 223kPa 7:02
136.4 C 223kPa 7:02
VENTING 7:02
AIR DRYING 7:54
CYCLE COMPLETE 7:56

Digital Signature #
2BC38600E7A6C9C1
