
Midmark M11 - v1.0.0
Total Cycles: 908

Name: RITTER 1
Sterilizer ID: V2178198
06/17/2021 06:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.7 F 0.0
2:00 77.7 F 0.0
4:00 110.6 F 0.0
6:00 146.5 F 0.0
8:00 152.4 F 0.1
10:00 171.3 F 0.2
12:00 185.2 F 1.3
14:00 197.9 F 3.9
16:00 212.3 F 6.4
18:00 223.0 F 9.6
20:00 232.6 F 13.2
22:00 231.8 F 8.3
24:00 242.2 F 12.8
26:00 239.6 F 10.3
28:00 237.4 F 8.8
30:00 248.8 F 14.5
32:00 259.5 F 20.3
34:00 269.4 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.7 F 28.8
2:00 271.9 F 28.1
3:00 271.8 F 28.4
4:00 272.7 F 28.7

Min 271.5 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.2 F 0.0
5:00 210.3 F 0.0
10:00 211.7 F 0.0
15:00 193.3 F 0.0
20:00 190.6 F 0.0
25:00 212.0 F 0.0
30:00 217.0 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 35:09
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:13:36
CYCLE COMPLETE

06/17/2021 07:57 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____