

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1870

Name: RITTER 1  
Sterilizer ID: V2178198  
04/26/2022 10:20 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 166.2 F 0.0  
2:00 163.3 F 2.2  
4:00 173.4 F 5.3  
6:00 204.9 F 10.0  
8:00 218.1 F 4.7  
10:00 230.4 F 9.0  
12:00 237.2 F 10.4  
14:00 240.1 F 11.3  
16:00 238.0 F 9.3  
18:00 251.6 F 16.0  
20:00 263.1 F 22.7  
22:00 272.4 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.4 F 28.8  
2:00 271.6 F 28.4  
3:00 272.7 F 28.8  
4:00 271.4 F 28.2

Min 271.4 F 28.1  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.9 F 0.0  
5:00 226.8 F 0.0  
10:00 228.2 F 0.0  
15:00 243.9 F 0.0  
20:00 231.9 F 0.0  
25:00 231.3 F 0.0  
30:00 224.7 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 22:11  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:00:18  
CYCLE COMPLETE

04/26/2022 11:25 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02