
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002137
9:43 10/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:07
PRESSURIZING 2:15
STERILIZING 3:18
136.4 C 223kPa 3:18
136.3 C 223kPa 3:48
136.2 C 221kPa 4:18
136.2 C 223kPa 4:48
136.3 C 224kPa 5:18
136.4 C 224kPa 5:48
136.5 C 223kPa 6:18
136.2 C 222kPa 6:48
136.2 C 222kPa 6:48
VENTING 6:48
AIR DRYING 7:43
CYCLE COMPLETE 37:43

Digital Signature #
0B2FCDB395AA6BA2
