
Midmark M11 - v1.0.0
Total Cycles: 1274

Name: RITTER 1
Sterilizer ID: V2178198
10/18/2021 11:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.9 F 0.0
2:00 197.3 F 2.6
4:00 207.8 F 6.9
6:00 221.7 F 12.0
8:00 225.3 F 6.6
10:00 236.4 F 11.5
12:00 235.3 F 8.8
14:00 241.1 F 10.5
16:00 247.4 F 13.9
18:00 259.5 F 20.5
20:00 270.4 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.2
1:00 272.8 F 28.9
2:00 271.7 F 28.1
3:00 272.7 F 28.8
4:00 271.6 F 28.1

Min 271.6 F 28.1
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.7 F 0.0
5:00 192.1 F 0.0
10:00 193.2 F 0.0
15:00 191.7 F 0.0
20:00 191.7 F 0.0
25:00 192.0 F 0.0
30:00 190.5 F 0.0
Drying Complete

FILLING: 1:50
HEATING: 20:43
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 00:59:54
CYCLE COMPLETE

10/18/2021 12:03 PM
ACCEPTED: X
REJECTED:
Operator: U01