
Midmark M11 - v1.0.0
Total Cycles: 1244

Name: RITTER 1
Sterilizer ID: V2178198
10/07/2021 03:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.4 F 0.0
2:00 179.9 F 3.4
4:00 200.9 F 7.4
6:00 221.5 F 12.6
8:00 227.7 F 7.4
10:00 240.8 F 12.7
12:00 238.1 F 9.9
14:00 236.3 F 8.4
16:00 249.5 F 14.8
18:00 261.6 F 21.6
20:00 272.4 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.0
1:00 272.6 F 28.8
2:00 271.7 F 28.1
3:00 271.6 F 28.5
4:00 272.7 F 28.8

Min 271.1 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.8 F 0.0
5:00 229.1 F 0.0
10:00 227.3 F 0.0
15:00 230.8 F 0.0
20:00 194.0 F 0.0
25:00 195.4 F 0.0
30:00 193.2 F 0.0
Drying Complete

FILLING: 1:45
HEATING: 20:28
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:22
CYCLE COMPLETE

10/07/2021 04:31 PM
ACCEPTED: X
REJECTED:
Operator: U02