
Midmark M11 - v1.0.5
Total Cycles: 772

Name: RITTER2
Sterilizer ID: V2313228
02/08/2022 02:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.1 F 0.0
2:00 198.6 F 5.2
4:00 222.5 F 12.0
6:00 231.8 F 9.1
8:00 237.9 F 9.7
10:00 240.0 F 10.2
12:00 252.0 F 16.7
14:00 265.4 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.3 F 29.1
2:00 272.5 F 29.7
3:00 271.9 F 28.8
4:00 272.8 F 29.6

Min 271.5 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.2 F 0.0
5:00 222.7 F 0.0
10:00 233.7 F 0.0
15:00 228.3 F 0.0
20:00 223.0 F 0.0
25:00 221.3 F 0.0
30:00 220.7 F 0.0
Drying Complete

FILLING: 2:55
HEATING: 15:34
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:55:41
CYCLE COMPLETE

02/08/2022 03:03 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____