

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002274  
9:04 18/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:30  
PRESSURIZING 3:32  
STERILIZING 4:37  
136.2 C 220kPa 4:37  
136.2 C 222kPa 5:07  
136.2 C 222kPa 5:37  
136.0 C 222kPa 6:07  
136.2 C 222kPa 6:37  
136.4 C 224kPa 7:07  
136.4 C 223kPa 7:37  
136.3 C 223kPa 8:07  
136.3 C 223kPa 8:07  
VENTING 8:07  
AIR DRYING 9:00  
CYCLE COMPLETE 12:24  
-----

Digital Signature #  
1D0AD09F904A9972  
-----