
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004406
10:25 21/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:31
PRESSURIZING 2:35
STERILIZING 3:21
136.0 C 220kPa 3:21
136.4 C 225kPa 3:51
136.2 C 221kPa 4:21
136.4 C 224kPa 4:51
136.4 C 224kPa 5:21
136.4 C 224kPa 5:51
136.4 C 224kPa 6:21
135.9 C 218kPa 6:51
135.9 C 218kPa 6:51
VENTING 6:51
AIR DRYING 7:45
CYCLE COMPLETE 10:40

Digital Signature #
383904E5B8CEB72D
