
Midmark M11 - v1.0.5
Total Cycles: 1508

Name: RITTER2
Sterilizer ID: V2313228
09/20/2022 11:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 130.9 F 0.0
2:00 146.0 F 1.2
4:00 159.2 F 4.4
6:00 177.3 F 9.2
8:00 206.2 F 6.8
10:00 219.9 F 8.2
12:00 233.5 F 13.2
14:00 235.6 F 10.5
16:00 235.4 F 8.7
18:00 249.0 F 15.7
20:00 261.2 F 22.6
22:00 272.3 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.4
1:00 272.7 F 30.1
2:00 271.9 F 29.3
3:00 272.3 F 30.1
4:00 272.7 F 29.9

Min 271.6 F 29.3
Max 272.8 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.4 F 0.0
5:00 192.9 F 0.0
10:00 204.4 F 0.0
15:00 211.0 F 0.0
20:00 212.9 F 0.0
25:00 212.9 F 0.0
30:00 215.0 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 22:30
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:01:03
CYCLE COMPLETE

09/20/2022 12:54 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____