
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001357
11:00 23/11/2022
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CYCLE ABORTED 0:05
STOP BUTTON PRESSED
NOT STERILE
END TIME 0:05

Digital Signature #
050BCDD3E92A493B
