

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1413

Name: RITTER2  
Sterilizer ID: V2313228  
08/25/2022 08:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 142.0 F 0.0  
2:00 142.6 F 0.0  
4:00 145.1 F 0.8  
6:00 177.8 F 3.3  
8:00 213.2 F 6.6  
10:00 217.9 F 11.3  
12:00 228.6 F 6.7  
14:00 236.2 F 11.9  
16:00 237.2 F 9.3  
18:00 238.8 F 9.8  
20:00 247.9 F 14.2  
22:00 259.5 F 20.7  
24:00 270.1 F 27.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.7  
1:00 272.3 F 29.0  
2:00 271.5 F 29.1  
3:00 272.0 F 28.9  
4:00 271.5 F 29.1

Min 271.1 F 28.5  
Max 272.6 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.4 F 0.0  
5:00 213.5 F 0.0  
10:00 223.7 F 0.0  
15:00 222.6 F 0.0  
20:00 226.8 F 0.0  
25:00 225.3 F 0.0  
30:00 220.3 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 24:55  
STERILIZING: 4:00  
VENTING: 3:00  
DRYING: 30:00  
TOTAL CYCLE: 01:02:48  
CYCLE COMPLETE

08/25/2022 10:00 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_