
Midmark M11 - v1.0.0
Total Cycles: 4566

Name: RITTER 1
Sterilizer ID: V2178198
05/21/2024 08:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.1 F 0.0
2:00 71.6 F 0.2
4:00 86.8 F 0.8
6:00 111.4 F 2.6
8:00 137.1 F 5.0
10:00 153.9 F 7.9
12:00 180.8 F 11.5
14:00 201.8 F 2.6
16:00 209.8 F 5.4
18:00 218.7 F 8.7
20:00 229.0 F 12.4
22:00 230.8 F 7.7
24:00 241.8 F 12.6
26:00 239.4 F 10.2
28:00 251.2 F 16.0
30:00 261.6 F 21.9
32:00 271.3 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.4 F 29.0
2:00 271.5 F 28.6
3:00 272.3 F 28.9
4:00 271.4 F 28.5

Min 271.4 F 28.3
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.6 F 0.0
5:00 206.7 F 0.0
10:00 206.2 F 0.0
15:00 212.8 F 0.0
20:00 218.7 F 0.0
25:00 231.9 F 0.0
30:00 227.9 F 0.0

Drying Complete

FILLING: 0:55
HEATING: 32:40
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:11:01
CYCLE COMPLETE

05/21/2024 09:32 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____