
Midmark M11 - v1.0.5
Total Cycles: 1702

Name: RITTER2
Sterilizer ID: V2313228
11/09/2022 08:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.4 F 0.0
2:00 78.7 F 0.0
4:00 82.2 F 0.0
6:00 98.4 F 0.0
8:00 152.4 F 0.4
10:00 186.7 F 0.9
12:00 207.6 F 3.8
14:00 218.3 F 7.2
16:00 228.8 F 11.2
18:00 230.2 F 6.7
20:00 241.5 F 12.1
22:00 239.2 F 10.3
24:00 237.2 F 8.9
26:00 251.3 F 15.9
28:00 263.0 F 22.7
30:00 272.3 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.2 F 28.8
2:00 272.3 F 29.2
3:00 271.7 F 28.4
4:00 272.6 F 29.0

Min 271.6 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.8 F 0.0
5:00 207.5 F 0.0
10:00 218.1 F 0.0
15:00 228.6 F 0.0
20:00 225.8 F 0.0
25:00 225.5 F 0.0
30:00 230.0 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 30:10
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:08:10
CYCLE COMPLETE

11/09/2022 09:29 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____