
Midmark M11 - v1.0.0
Total Cycles: 3165

Name: RITTER 1
Sterilizer ID: V2178198
04/27/2023 07:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.2 F 0.0
2:00 79.9 F 0.0
4:00 94.6 F 0.0
6:00 113.1 F 0.0
8:00 145.4 F 1.9
10:00 173.4 F 3.7
12:00 188.4 F 6.3
14:00 204.0 F 10.0
16:00 215.1 F 10.1
18:00 223.3 F 6.4
20:00 233.1 F 10.4
22:00 235.4 F 9.0
24:00 242.3 F 11.8
26:00 239.5 F 9.9
28:00 252.1 F 16.1
30:00 262.9 F 22.4
32:00 272.7 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.2 F 28.3
2:00 272.5 F 28.9
3:00 271.5 F 28.0
4:00 272.4 F 28.5

Min 271.4 F 27.9
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.8 F 0.0
5:00 202.3 F 0.0
10:00 203.8 F 0.0
15:00 204.8 F 0.0
20:00 205.0 F 0.0
25:00 208.2 F 0.0
30:00 206.8 F 0.0

Drying Complete

FILLING: 0:47
HEATING: 32:17
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:10:26
CYCLE COMPLETE

04/27/2023 08:22 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____