
Midmark M11 - v1.0.0
Total Cycles: 2657

Name: RITTER 1
Sterilizer ID: V2178198
12/08/2022 08:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 79.9 F 0.0
2:00 84.1 F 0.0
4:00 121.8 F 0.0
6:00 127.5 F 0.9
8:00 151.3 F 2.4
10:00 177.4 F 4.8
12:00 194.7 F 8.4
14:00 210.9 F 12.1
16:00 219.8 F 5.2
18:00 229.7 F 9.4
20:00 237.4 F 10.9
22:00 239.4 F 10.8
24:00 235.6 F 8.4
26:00 249.6 F 14.9
28:00 261.1 F 21.3
30:00 271.4 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.4 F 28.7
2:00 271.8 F 28.5
3:00 272.3 F 28.6
4:00 271.8 F 28.6

Min 271.5 F 28.1
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.9 F 0.0
5:00 216.5 F 0.0
10:00 220.8 F 0.0
15:00 224.6 F 0.0
20:00 232.0 F 0.0
25:00 222.9 F 0.0
30:00 223.6 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 30:39
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:08:31
CYCLE COMPLETE

12/08/2022 09:59 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____