
Midmark M11 - v1.0.0
Total Cycles: 2429

Name: RITTER 1
Sterilizer ID: V2178198
10/13/2022 12:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.1 F 0.0
2:00 177.7 F 4.0
4:00 196.0 F 8.6
6:00 218.3 F 9.7
8:00 231.8 F 8.3
10:00 240.6 F 12.4
12:00 241.1 F 11.4
14:00 238.9 F 10.0
16:00 253.1 F 16.9
18:00 265.1 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.5 F 28.8
2:00 271.8 F 28.8
3:00 272.1 F 28.6
4:00 271.7 F 28.8

Min 271.2 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.2 F 0.0
5:00 233.7 F 0.0
10:00 211.0 F 0.0
15:00 219.9 F 0.0
20:00 208.3 F 0.0
25:00 207.1 F 0.0
30:00 210.7 F 0.0
Drying Complete

FILLING: 2:12
HEATING: 19:45
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:11
CYCLE COMPLETE

10/13/2022 01:06 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____