
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001484
9:46 26/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:05
STERILIZING 4:14
136.4 C 225kPa 4:14
136.4 C 225kPa 4:44
136.4 C 225kPa 5:14
136.4 C 225kPa 5:44
136.3 C 224kPa 6:14
136.3 C 224kPa 6:44
136.2 C 222kPa 7:14
136.3 C 225kPa 7:44
136.3 C 225kPa 7:44
VENTING 7:44
AIR DRYING 8:37
CYCLE COMPLETE 8:44

Digital Signature #
342102E5E8E8CF80
