
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000706
7:19 20/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:24
PRESSURIZING 1:28
STERILIZING 2:33
136.4 C 230kPa 2:33
136.3 C 231kPa 3:03
136.4 C 232kPa 3:33
136.4 C 230kPa 4:03
136.3 C 231kPa 4:33
135.9 C 228kPa 5:03
136.4 C 231kPa 5:33
136.3 C 231kPa 6:03
136.3 C 231kPa 6:03
VENTING 6:03
AIR DRYING 6:58
CYCLE COMPLETE 7:03

Digital Signature #
7D6379FD6B76430C
