
Midmark M11 - v1.0.5
Total Cycles: 3030

Name: RITTER2
Sterilizer ID: V2313228
10/19/2023 10:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.0 F 0.0
2:00 179.4 F 2.9
4:00 192.6 F 8.0
6:00 213.9 F 12.3
8:00 225.9 F 9.0
10:00 234.4 F 9.7
12:00 243.0 F 13.1
14:00 243.5 F 12.4
16:00 256.8 F 19.5
18:00 268.4 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.6 F 29.7
2:00 271.8 F 28.9
3:00 272.0 F 29.5
4:00 272.5 F 29.5

Min 271.3 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.8 F 0.0
5:00 231.9 F 0.0
10:00 225.2 F 0.0
15:00 213.2 F 0.0
20:00 210.8 F 0.0
25:00 217.6 F 0.0
30:00 217.9 F 0.0

Drying Complete

FILLING: 1:46
HEATING: 19:12
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:12
CYCLE COMPLETE

10/19/2023 11:24 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____