
Midmark M11 - v1.0.5
Total Cycles: 295

Name: RITTER2
Sterilizer ID: V2313228
08/20/2021 10:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.9 F 0.0
2:00 173.4 F 2.4
4:00 184.5 F 6.2
6:00 205.4 F 10.8
8:00 219.9 F 5.8
10:00 230.7 F 11.3
12:00 233.8 F 9.1
14:00 237.9 F 9.8
16:00 249.1 F 15.5
18:00 261.0 F 22.3
20:00 271.4 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.4
1:00 272.4 F 29.7
2:00 271.7 F 29.8
3:00 272.0 F 29.6
4:00 271.5 F 29.8

Min 271.2 F 29.2
Max 272.4 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.2 F 0.0
5:00 234.0 F 0.0
10:00 225.9 F 0.0
15:00 215.7 F 0.0
20:00 193.7 F 0.0
25:00 187.3 F 0.0
30:00 184.8 F 0.0
Drying Complete

FILLING: 1:54
HEATING: 20:35
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:44
CYCLE COMPLETE

08/20/2021 11:00 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____