
Midmark M11 - v1.0.0
Total Cycles: 4833

Name: RITTER 1
Sterilizer ID: V2178198
08/19/2024 04:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.2 F 0.0
2:00 163.6 F 2.0
4:00 191.8 F 6.8
6:00 214.5 F 12.7
8:00 227.5 F 7.9
10:00 238.9 F 11.2
12:00 241.1 F 11.4
14:00 238.8 F 9.9
16:00 253.8 F 17.3
18:00 266.4 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.3
1:00 272.6 F 29.0
2:00 271.4 F 28.2
3:00 272.4 F 28.9
4:00 271.4 F 28.3

Min 271.2 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.1 F 0.0
5:00 227.9 F 0.0
10:00 209.7 F 0.0
15:00 211.5 F 0.0
20:00 220.9 F 0.0
25:00 243.3 F 0.0
30:00 215.7 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 19:20
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:51
CYCLE COMPLETE

08/19/2024 05:42 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____