
Midmark M11 - v1.0.0
Total Cycles: 2516

Name: RITTER 1
Sterilizer ID: V2178198
11/07/2022 08:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.3 F 0.0
2:00 74.7 F 0.0
4:00 91.8 F 0.0
6:00 135.4 F 0.0
8:00 153.3 F 0.1
10:00 178.2 F 3.1
12:00 198.8 F 6.6
14:00 217.3 F 10.0
16:00 226.1 F 11.8
18:00 226.6 F 7.6
20:00 236.2 F 11.9
22:00 233.5 F 8.4
24:00 242.9 F 11.6
26:00 244.5 F 12.5
28:00 256.2 F 18.7
30:00 266.8 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.0 F 28.5
2:00 272.4 F 28.8
3:00 272.7 F 29.2
4:00 271.8 F 28.7

Min 271.5 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.8 F 0.0
5:00 204.5 F 0.0
10:00 191.9 F 0.0
15:00 196.0 F 0.0
20:00 193.2 F 0.0
25:00 205.9 F 0.0
30:00 220.5 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 31:29
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:09:25
CYCLE COMPLETE

11/07/2022 09:54 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____