
Midmark M11 - v1.0.0
Total Cycles: 2826

Name: RITTER 1
Sterilizer ID: V2178198
01/30/2023 01:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.6 F 0.0
2:00 160.8 F 3.6
4:00 180.4 F 7.5
6:00 198.9 F 12.2
8:00 216.3 F 5.5
10:00 228.7 F 10.4
12:00 231.2 F 7.6
14:00 242.5 F 12.7
16:00 240.9 F 11.0
18:00 254.1 F 17.6
20:00 265.5 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.5 F 28.9
2:00 272.0 F 29.0
3:00 272.2 F 28.8
4:00 272.0 F 29.0

Min 271.6 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.6 F 0.0
5:00 208.1 F 0.0
10:00 224.6 F 0.0
15:00 229.0 F 0.0
20:00 229.2 F 0.0
25:00 227.5 F 0.0
30:00 225.7 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 21:40
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:00:02
CYCLE COMPLETE

01/30/2023 02:48 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____