
Midmark M11 - v1.0.5
Total Cycles: 2619

Name: RITTER2
Sterilizer ID: V2313228
07/06/2023 01:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.2 F 0.0
2:00 180.4 F 3.1
4:00 186.5 F 7.6
6:00 216.9 F 13.2
8:00 227.3 F 8.4
10:00 239.3 F 11.5
12:00 240.0 F 11.8
14:00 239.6 F 10.7
16:00 253.1 F 17.5
18:00 264.5 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.0 F 29.3
2:00 272.1 F 29.7
3:00 271.3 F 29.1
4:00 272.1 F 29.6

Min 271.3 F 29.0
Max 272.3 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.2 F 0.0
5:00 216.9 F 0.0
10:00 236.4 F 0.0
15:00 234.8 F 0.0
20:00 226.1 F 0.0
25:00 222.5 F 0.0
30:00 221.2 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 19:50
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:39
CYCLE COMPLETE

07/06/2023 02:38 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____